

# EC Declaration of Conformity

In Accordance with EN ISO/IEC 17050-1:2010

Manufacturer: Nordson EFD LLC  
Address: 40 Catamore Boulevard  
East Providence, RI 02914 USA Type of  
Equipment: Fluid Dispenser(s)  
Product Name: Performus Dispenser Models X15 / X100  
Model & Part Number: ☐ Performus X15, 7363257  
☐ Performus X100, 7363256  
Manufacture Date: \_\_\_\_\_ Serial Number: \_\_\_\_\_

The above listed product(s) have been evaluated for conformity to the essential requirements of the below-listed directives:

WEEE Directive	2012/19/EU
'RoHS Directive'	2011/65/EU
'EMC Directive'	2014/30/EU
Low Voltage	2014/35/EU

With reference to the follow harmonized standards:

## Emissions Standards:

EN61000-6-3:2007+A1:2011	Conducted Emissions at AC Mains
EN61000-6-3:2007+A1:2011	Radiated Emissions
Subpart B-15.107	Conducted Emissions
Subpart B-15.109	Radiated Emissions
EN61000-3-2:2006 +A1:2009+ A2:2009	Harmonic Current Emissions
EN61000-3-3:2008	Voltage Fluctuations Emissions

## Immunity Standards:

EN61000-4-2:1995+A1:1998+ A2:2010	Electrostatic Discharge (ESD) Immunity
EN61000-4-3:2006+A1:2008+ A2:2010	RF Electromagnetic Field Immunity
EN61000-4-4:2004+A1:2010	Fast Transient Immunity
EN61000-4-5:2006	Surges Immunity
EN61000-4-6:2013	Power-Frequency Magnetic Field Immunity
EN61000-4-11:2004	Voltage Dips/Voltage Interruptions Immunity

## Safety Standards:

IEC 61010-1:2010 Third Edition  
Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use,  
Part 1: General Requirements

Nordson EFD LLC

*Wayne Burkett*

Plant Manager  
Place: East Providence, RI USA  
Date: January 24, 2018

Authorized Representative:  
European Technical Manager EFD  
International  
Unit 14, Apex Business Centre  
Boscombe Road  
Dunstable, Bedfordshire LU5 4SB  
England